

Notice of Allowability

Application No.

10/630,468

Examiner

Jon-Eric C. Morales

Applicant(s)

HOLMSTROM, NILS

Art Unit

3766

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 12/26/2007.
2. ☒ The allowed claim(s) is/are 1, 3, 5-13, 15, 17-28, 30 and 32-39.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☒ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____

DETAILED ACTION

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

The application has been amended as follows:

1. (Currently Amended) An implantable heart monitoring device comprising:

a control circuit;

a first sensor adapted to be disposed in a coronary sinus region of a heart to sense blood oxygen in said coronary sinus region, and supplying a first sensor signal representing a blood oxygen level to the control circuit;

at least one further sensor adapted to interact with the heart to sense activity of the heart, said at least one further sensor supplying a further sensor signal representing said activity to the control circuit; and

said control circuit, from said further sensor signal, identifying a first portion in a diastolic portion, and a second portion in a systolic portion, of a same heart cycle of the heart, and from said first sensor signal, determining a first value related to said blood oxygen level that occurs during said first portion and identifying said first value as having occurred during said first portion and determining a second value related to said blood constituent oxygen level that occurs during said second portion and identifying said second value as having occurred during said second portion, and said control circuit emitting an output signal indicative of functioning of the heart dependent on a relation between said first value and said second value.

Art Unit: 3766

2. The following is an examiner's statement of reasons for allowance: In regards to independent claims 1, 13 and 28 a implantable heart monitoring device includes a control circuit from sensor signal to identify a first portion n a diastolic portion, and a second portion in a systolic portion of the same heart cycle and from first sensor signal determining a first value related to said blood oxygen level and determining a second value related to said blood oxygen level has not been disclosed or suggested in the prior art when combined with the rest of the limitation of claims 1 , 13, and 28.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

3. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jon-Eric C. Morales whose telephone number is 571-272-3107. The examiner can normally be reached on Monday through Friday from 8am - 5pm EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Carl H. Layno can be reached on 571-272-4949. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3766

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Jon-Eric Morales
/Jon-Eric Morales/
Patent Examiner
Art Unit 3766

JEM

Carl Layno
Supervisory Patent Examiner
Art Unit 3766

Carl H. Layno
CARL LAYNO
PRIMARY EXAMINER